

Triiodothyronine (T3) ELISA (Mouse/Rat) Kit

Name

Triiodothyronine (T3) ELISA (Mouse/Rat) Kit

INTENDED USE

The Calbiotech Mouse/Rat Triiodothyronine (T3) ELISA Kit is intended for the quantitative measurement of Triiodothyronine (T3) in mouse/rat serum or plasma. For Research Use Only. For professional use only. Not for use in diagnostic procedures.

SUMMARY AND EXPLANATION

Triiodothyronine (T3) is a thyroid hormone influencing metabolism, thermogenesis, and development. This assay measures T3 in rodent samples for research on thyroid hormone synthesis, metabolism, and energy regulation in animal models.

MATERIALS PROVIDED	96 TESTS
Microwells coated with T3 Monoclonal Ab	12x8x1
T3 Standard: 7 vials (ready to use)	0.25 mL
T3 Controls: 2 vials (ready to use)	0.25 mL
T3 Enzyme Conjugate concentrate: 1 vial	1.5 mL
Assay diluent: (ready to use)	12 mL
TMB Substrate: 1 bottle (ready to use)	12 mL
Stop Solution: 1 bottle (ready to use)	12 mL
20X Wash Concentrate: 1 bottle	25 mL

MATERIALS NOT PROVIDED

Distilled or deionized water

Precision pipettes

Disposable pipette tips

ELISA reader capable of reading absorbance at 450nm

STORAGE AND STABILITY

Absorbance paper or paper towel

Graph paper

1. Store the kit at 2 – 8 C.

Keep microwells sealed in a dry bag with desiccants.

The reagents are stable until expiration of the kit.

Do not expose test reagents to heat, sun, or strong light.

WARNINGS AND PRECAUTIONS

Potential biohazardous materials:

The standards contain human source components which have been tested and found non-reactive for hepatitis B surface antigen as well as HIV antibody with FDA licensed reagents. However, as there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent, these reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories." 1984.

Do not pipette by mouth. Do not smoke, eat, or drink in the areas in which specimens or kit reagents are handled.

The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed.

It is recommended that standards, control and serum samples be run in duplicate

Optimal results will be obtained by strict adherence to this protocol. Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential. Any deviation from this may yield invalid data.

SPECIMEN COLLECTION HANDLING

This assay is designed for use with mouse/rat serum or plasma samples obtained in compliance with applicable laws, regulations, and institutional policies. Handle and store samples using procedures appropriate for research use. Samples may be stored refrigerated (2–8 °C) for up to seven days, or frozen (–20 °C or below) for up to six months. Avoid repetitive freeze–thaw cycles.

REAGENT PREPARATION

T3-enzyme Conjugate Solution

Dilute the T3-enzyme conjugate 1:11 with assay diluent in a suitable container. For example, dilute 160µl of conjugate with 1.6ml of assay diluent for 16 wells (A slight excess of solution is made). This reagent should be used within twenty-four hours for maximum performance of the assay. Store at 2-8°C. General Formula:

Amount of Buffer required = Number of wells * 0.1

Quantity of T3-Enzyme necessary = # of wells * 0.01

i.e. = 16 x 0.1 = 1.6ml for Assay diluent

16 x 0.01 = 0.16ml (160µl) for T3 enzyme conjugate

Wash Buffer

Prepare 1X Wash buffer by adding the contents of the bottle (25 ml, 20X) to 475 ml of distilled or deionized water. Store at room temperature (20-25°C).

ASSAY PROCEDURE

Before proceeding with the assay, bring all reagents, serum references and controls to room temperature (20-25°C).

Format the microplate wells for each serum reference, control and patient specimen to be assayed in duplicate. Replace any unused microwell strips back into the aluminum bag, seal and store at 2-8°C.

Pipette 25µl of the appropriate serum reference, control or specimen into the assigned well.

Add 100µl of working T3-enzyme conjugate solution to all wells (see Reagent Preparation Section).

Cover the plate and Incubate for 60 minutes at room temperature with shaking at 650rpm.

Remove liquid from all wells. Wash wells three times with 300 of 1X wash buffer (see Reagent Preparation Section). Blot on absorbent paper towels.

Add 100µl of TMB substrate solution to all wells

Cover the plate and Incubate at room temperature for fifteen (15) minutes.

Add 50µl of stop solution to each well and gently mix for 15-20 seconds.

CALCULATION OF RESULTS

Read the absorbance on ELISA Reader of each well at 450nm within 15 minutes after adding the stop solution.

The standard curve is constructed as follows:

Check T3 standard value on each standard vial. This value might vary from lot to lot. Make sure you check the value on every kit. See example of the standard attached.

To construct the standard curve, plot the absorbance for T3 standards (vertical axis) versus T3 standard concentrations (horizontal axis) on a linear graph paper. Draw the best curve through the points.

Read the absorbance for controls and each unknown sample from the curve. Record the value for each control or unknown sample.

Example of a standard Curve

	OD 450 nm	Conc. ng/mL
Std 1	2.404	0
Std 2	2.238	0.25
Std 3	2.001	0.5
Std 4	1.714	1
Std 5	1.147	2.5
Std 6	0.793	5
Std 7	0.642	7.5

LIMITATIONS OF THE TEST

Do not use sodium azide as preservative. Sodium azide inhibits HRP enzyme activities.

REFERENCES

Agharanya JC. Clinical usefulness of ELISA technique in the assessment of thyroid function. West Afr J Med 1990;9(4):258-63.

Baumgartner-Parzer SM; Wagner L; Reining G; Sexl V; Nowotny P; Muller M; Brunner M; Waldh  usl W. Increase by triiodothyronine of endothelin-1, fibronectin and von Willebrand factor in cultured endothelial cells. J Endocrinol 1997;154(2):231-9.

Maes M; Mommen K; Hendrickx D; Peeters D; D'Hondt P; Ranjan R; De Meyer F; Scharp  e S Components of biological variation, including seasonality, in blood concentrations of TSH, TT3, FT4, PRL, cortisol and testosterone in healthy volunteers. Clin Endocrinol (Oxf)

1997; 46(5):587-98.

Santini F; Chiovato L; Bartalena L; Lapi P; Palla R; Panichi V; Velluzzi F; Grasso L; Chopra IJ; Martino E; Pinchera A
Study of serum 3,5,3'-triiodothyronine sulfate
concentration in patients with systemic non-thyroidal
illness. Eur J Endocrinol 1996;134(1):45-9